

# CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

DATE OF ISSUE: 16 March 2016

CERTIFICATE No: 389258




**Lambda**  
CALIBRATION LTD

Units 11-13  
Chorley Central Business Park  
Stump Lane, Chorley  
Lancashire PR6 0BL  
Tel: 0845 2411533  
Fax: 0845 2411544

Page 1 of 2

APPROVED SIGNATAORY

  
A Kelly D Pilkington  
D Whalley C Reed R Armitage

**Customer:** DJB Labcare Ltd  
**Address:** 20 Howard Way, Interchange Park,  
Milton Keynes  
MK16 9QS

**Item Number:** 12078435 (4046)  
**Description:** Tachometer  
**Model/Range:** AT-6  
**Manufacturer:** Standard  
**Date of Cal:** 16 Mar 2016  
**Calibrated by:** Mohammed Abid  
**Procedure Name:** Standard, Digital Tachometer, AT-6  
**Rev/Basis:** 01:E-1000  
**Temp/Humidity:** 20.0°C ± 2°C <80%rh

The Results on the following pages are: As Found

All Measurements are Traceable to National Standards.

Note 1: The unit under test was calibrated using a multifunction calibrator.

Note 2: Where the reported value lies within the specified tolerances then this will be indicated by the word "PASS", if outside then by the word "FAIL".

Note 3: Values quoted in the "UUT Indicated Value" column are not necessarily quoted to the same resolution as the actual displayed value on the UUT.

Note 4: Any supplied test leads have been checked as part of the Visual/Operational test but have not been used during calibration.

**Engineers' Notes:**

**Standard(s) Used:** LMMC-02 / LMMC-04 / LMMC-10 / LMMC-14

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

UKAS ACCREDITED CALIBRATION LABORATORY No: 0495

CERTIFICATE No:

389258

Page 2 of 2

Parameter	UUT Range	Applied Value	UUT Reading	Acceptance Limits Low	Acceptance Limits High	Pass/Fail
Visual/Operational Test						
Result of Operator Evaluation						PASS
Various frequencies were applied from a multi-function calibrator to the tachometer optical calibration box, LSTR-06, the output of which was measured with the UUT. The results were as follows:						
RPM		600.0rpm	600.0	599.6	600.4	PASS
		1000rpm	1000	999	1002	PASS
		2500rpm	2500	2498	2502	PASS
		5000rpm	5003	4997	5004	PASS
		7500rpm	7500	7495	7505	PASS
		10000rpm	10000	9994	10006	PASS
		14000rpm	14000	13992	14008	PASS
		18000rpm	18001	17990	18010	PASS
		20000rpm	19999	19989	20011	PASS
		25000rpm	24997	24987	25014	PASS
		30000rpm	30002	29984	30016	PASS
		40000rpm	39998	39979	40021	PASS
		50000rpm	50001	49974	50026	PASS
		60000rpm	60003	59969	60031	PASS
		70000rpm	70001	69964	70036	PASS
		80000rpm	80003	79959	80041	PASS
		90000rpm	90006	89954	90046	PASS

End of Calibration Data

Estimated Uncertainty of Measurement:

-----  
Frequency

-----  
0.2Hz to 250MHz: +/- (0.29ppm + 2 LSD)