



Certificate of Calibration

Certificate number: 180817

Equipment: Fastime 3 Stopwatch
Serial number: 018

Date: 19/08/2017
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	04:59.92
10:00.00	10:00.06
1:00:00.00	1:00:00.21

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith