



Certificate of Calibration

Certificate number: 120218

Equipment: Fastime 3 Stopwatch
Serial number: 012

Date: 09/02/18
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	05:00.07
10:00.00	10:00.08
60:00.00	60:00.09

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam